
Midmark M11 - v1.0.0
Total Cycles: 4527

Name: RITTER 1
Sterilizer ID: V2178198
05/06/2024 12:48 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 95.0 F 0.0
2:00 96.1 F 0.1
4:00 113.8 F 1.8
6:00 145.0 F 4.2
8:00 167.4 F 6.9
10:00 188.1 F 9.9
12:00 212.0 F 9.2
14:00 212.7 F 5.1
16:00 223.3 F 8.6
18:00 232.9 F 12.3
20:00 231.5 F 8.3
22:00 242.5 F 13.0
24:00 240.6 F 10.9
26:00 251.8 F 16.4
28:00 261.9 F 22.1
30:00 271.5 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.4
1:00 272.3 F 29.2
2:00 272.2 F 28.9
3:00 271.4 F 28.3
4:00 271.6 F 28.8

Min 271.2 F 28.3
Max 272.5 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.8 F 0.0
5:00 213.2 F 0.0
10:00 207.8 F 0.0
15:00 221.4 F 0.0
20:00 233.8 F 0.0
25:00 227.9 F 0.0
30:00 228.2 F 0.0
Drying Complete

FILLING: 1:20
HEATING: 30:44
STERILIZING: 4:00
VENTING: 3:51
DRYING: 30:00
TOTAL CYCLE: 01:09:55
CYCLE COMPLETE

05/06/2024 01:57 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____