
Midmark M11 - v1.0.5
Total Cycles: 3242

Name: RITTER2
Sterilizer ID: V2313228
12/20/2023 06:28 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 174.9 F 0.0
2:00 174.7 F 3.9
4:00 201.6 F 10.4
6:00 220.0 F 6.0
8:00 236.6 F 12.6
10:00 239.4 F 11.0
12:00 239.9 F 10.4
14:00 255.3 F 18.2
16:00 268.3 F 26.3

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 28.5
1:00 272.5 F 28.9
2:00 272.8 F 29.2
3:00 272.2 F 29.0
4:00 271.9 F 28.4

Min 271.6 F 28.3
Max 272.8 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.6 F 0.0
5:00 223.4 F 0.0
10:00 235.9 F 0.0
15:00 229.2 F 0.0
20:00 223.3 F 0.0
25:00 223.6 F 0.0
30:00 224.0 F 0.0
Drying Complete

FILLING: 2:06
HEATING: 16:57
STERILIZING: 4:00
VENTING: 3:02
DRYING: 30:00
TOTAL CYCLE: 00:56:05
CYCLE COMPLETE

12/20/2023 07:24 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____