

-----  
Midmark M11 - v1.0.0  
Total Cycles: 877

Name: RITTER 1  
Sterilizer ID: V2178198  
06/09/2021 10:28 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 148.0 F 0.0  
2:00 152.5 F 1.5  
4:00 165.0 F 4.3  
6:00 191.4 F 7.6  
8:00 206.8 F 12.2  
10:00 216.6 F 5.1  
12:00 228.9 F 9.6  
14:00 234.9 F 9.5  
16:00 241.0 F 11.7  
18:00 239.2 F 10.0  
20:00 251.9 F 16.2  
22:00 263.0 F 22.7  
24:00 272.6 F 28.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 28.3  
1:00 272.8 F 29.0  
2:00 272.0 F 28.3  
3:00 272.5 F 28.9  
4:00 272.7 F 28.8

Min 271.7 F 28.2  
Max 272.9 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 207.9 F 0.0  
5:00 198.2 F 0.0  
10:00 199.4 F 0.0  
15:00 198.2 F 0.0  
20:00 200.0 F 0.0  
25:00 203.3 F 0.0  
30:00 212.0 F 0.0  
Drying Complete

FILLING: 1:00  
HEATING: 24:19  
STERILIZING: 4:00  
VENTING: 3:26  
DRYING: 30:00  
TOTAL CYCLE: 01:02:46  
CYCLE COMPLETE

06/09/2021 11:30 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_