
Midmark M11 - v1.0.5
Total Cycles: 3192

Name: RITTER2
Sterilizer ID: V2313228
12/06/2023 03:24 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 137.3 F 0.0
2:00 141.6 F 1.5
4:00 167.7 F 4.6
6:00 187.9 F 8.4
8:00 206.5 F 12.6
10:00 216.3 F 7.0
12:00 227.4 F 11.8
14:00 234.4 F 8.7
16:00 239.9 F 10.7
18:00 246.3 F 13.6
20:00 258.3 F 20.2
22:00 269.2 F 27.2

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.4
1:00 272.8 F 30.1
2:00 271.3 F 29.3
3:00 272.4 F 30.0
4:00 271.4 F 29.3

Min 271.1 F 29.3
Max 272.9 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 177.2 F 0.0
5:00 239.5 F 0.0
10:00 203.9 F 0.0
15:00 179.4 F 0.0
20:00 176.2 F 0.0
25:00 173.9 F 0.0
30:00 172.7 F 0.0
Drying Complete

FILLING: 1:07
HEATING: 23:42
STERILIZING: 4:00
VENTING: 3:35
DRYING: 30:00
TOTAL CYCLE: 01:02:25
CYCLE COMPLETE

12/06/2023 04:26 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____