
Midmark M11 - v1.0.0
Total Cycles: 2859

Name: RITTER 1
Sterilizer ID: V2178198
02/06/2023 11:15 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 207.2 F 0.0
2:00 202.0 F 3.8
4:00 199.2 F 8.4
6:00 219.1 F 9.0
8:00 230.0 F 9.0
10:00 238.5 F 10.0
12:00 241.3 F 11.9
14:00 240.1 F 10.3
16:00 253.8 F 17.2
18:00 265.6 F 24.3

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.3
1:00 272.8 F 29.2
2:00 271.9 F 28.5
3:00 271.8 F 28.7
4:00 272.7 F 29.1

Min 271.3 F 28.2
Max 272.8 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.7 F 0.0
5:00 231.0 F 0.0
10:00 232.6 F 0.0
15:00 228.2 F 0.0
20:00 227.9 F 0.0
25:00 228.6 F 0.0
30:00 227.9 F 0.0
Drying Complete

FILLING: 1:45
HEATING: 19:46
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:58:40
CYCLE COMPLETE

02/06/2023 12:13 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____