
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002646
13:33 02/04/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:00
PRESSURIZING 3:05
STERILIZING 4:19
136.0 C 220kPa 4:19
136.2 C 223kPa 4:49
136.4 C 225kPa 5:19
136.1 C 222kPa 5:49
136.4 C 224kPa 6:19
136.1 C 222kPa 6:49
136.1 C 222kPa 7:19
136.4 C 225kPa 7:49
136.4 C 225kPa 7:49
VENTING 7:49
AIR DRYING 8:41
CYCLE COMPLETE 38:41

Digital Signature #
8E753C8A3820354F
