
Midmark M11 - v1.0.5
Total Cycles: 3477

Name: RITTER2
Sterilizer ID: V2313228
03/26/2024 11:14 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 108.7 F 0.0
2:00 117.0 F 0.0
4:00 118.1 F 0.1
6:00 146.0 F 2.3
8:00 168.0 F 5.6
10:00 190.4 F 9.2
12:00 201.9 F 13.1
14:00 217.2 F 6.8
16:00 227.7 F 11.1
18:00 229.8 F 7.7
20:00 242.6 F 13.2
22:00 241.5 F 11.8
24:00 253.2 F 17.7
26:00 263.8 F 24.0
28:00 272.5 F 29.6

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.2
1:00 271.5 F 29.7
2:00 272.7 F 29.8
3:00 271.5 F 29.2
4:00 271.4 F 29.7

Min 271.1 F 29.1
Max 272.8 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.9 F 0.0
5:00 225.4 F 0.0
10:00 230.0 F 0.0
15:00 230.4 F 0.0
20:00 221.9 F 0.0
25:00 225.3 F 0.0
30:00 227.5 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 28:07
STERILIZING: 4:00
VENTING: 3:37
DRYING: 30:00
TOTAL CYCLE: 01:06:38
CYCLE COMPLETE

03/26/2024 12:20 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____