
Midmark M11 - v1.0.5
Total Cycles: 1761

Name: RITTER2
Sterilizer ID: V2313228
11/22/2022 01:07 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 214.3 F 0.0
2:00 205.3 F 4.2
4:00 211.0 F 10.6
6:00 225.6 F 7.1
8:00 239.7 F 13.4
10:00 241.6 F 12.2
12:00 242.4 F 11.9
14:00 256.7 F 19.3
16:00 268.7 F 27.0

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.9
1:00 272.5 F 29.5
2:00 271.5 F 28.8
3:00 272.5 F 29.7
4:00 272.1 F 29.2

Min 271.4 F 28.8
Max 272.6 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.1 F 0.0
5:00 220.5 F 0.0
10:00 229.0 F 0.0
15:00 245.6 F 0.0
20:00 234.8 F 0.0
25:00 228.4 F 0.0
30:00 210.8 F 0.0
Drying Complete

FILLING: 2:50
HEATING: 17:06
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:57:08
CYCLE COMPLETE

11/22/2022 02:04 PM
ACCEPTED:
REJECTED:

Reason: _____
Operator: _____