
Midmark M11 - v1.0.5
Total Cycles: 3800

Name: RITTER2
Sterilizer ID: V2313228
07/19/2024 09:41 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 138.9 F 0.0
2:00 140.5 F 1.6
4:00 151.1 F 4.4
6:00 174.0 F 7.9
8:00 193.0 F 12.0
10:00 209.2 F 5.4
12:00 220.2 F 9.9
14:00 230.0 F 9.1
16:00 236.0 F 11.6
18:00 237.5 F 9.8
20:00 250.1 F 15.9
22:00 261.1 F 22.1
24:00 271.0 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.5
1:00 272.6 F 29.8
2:00 272.1 F 30.0
3:00 272.2 F 29.6
4:00 272.1 F 30.1

Min 271.5 F 29.4
Max 272.8 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 182.7 F 0.0
5:00 210.1 F 0.0
10:00 194.3 F 0.0
15:00 191.3 F 0.0
20:00 188.6 F 0.0
25:00 185.5 F 0.0
30:00 185.3 F 0.0
Drying Complete

FILLING: 0:58
HEATING: 25:31
STERILIZING: 4:00
VENTING: 3:44
DRYING: 30:00
TOTAL CYCLE: 01:04:13
CYCLE COMPLETE

07/19/2024 10:45 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____