

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1811

Name: RITTER 1  
Sterilizer ID: V2178198  
04/11/2022 01:09 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 210.6 F 0.0  
2:00 207.7 F 4.5  
4:00 213.7 F 10.7  
6:00 224.6 F 6.1  
8:00 236.2 F 11.9  
10:00 236.6 F 9.5  
12:00 239.2 F 8.9  
14:00 249.1 F 14.8  
16:00 261.2 F 21.6  
18:00 272.4 F 29.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.4  
1:00 272.6 F 29.1  
2:00 271.6 F 28.3  
3:00 272.3 F 29.2  
4:00 272.3 F 28.7  
  
Min 271.3 F 28.3  
Max 272.6 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 211.6 F 0.0  
5:00 233.2 F 0.0  
10:00 227.3 F 0.0  
15:00 231.5 F 0.0  
20:00 230.2 F 0.0  
25:00 229.5 F 0.0  
30:00 226.6 F 0.0  
Drying Complete

FILLING: 2:36  
HEATING: 18:28  
STERILIZING: 4:00  
VENTING: 3:09  
DRYING: 30:00  
TOTAL CYCLE: 00:58:12  
CYCLE COMPLETE

04/11/2022 02:07 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_