
Midmark M11 - v1.0.5
Total Cycles: 446

Name: RITTER2
Sterilizer ID: V2313228
10/13/2021 12:46 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 196.5 F 0.0
2:00 188.4 F 3.9
4:00 212.6 F 10.3
6:00 225.0 F 7.4
8:00 238.3 F 11.9
10:00 242.9 F 12.8
12:00 243.9 F 12.6
14:00 258.0 F 20.1
16:00 270.2 F 27.9

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.8
1:00 272.0 F 28.9
2:00 272.4 F 29.4
3:00 271.9 F 29.5
4:00 271.7 F 28.7

Min 271.4 F 28.7
Max 272.8 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 214.6 F 0.0
5:00 227.7 F 0.0
10:00 227.5 F 0.0
15:00 213.0 F 0.0
20:00 203.7 F 0.0
Dry Cycle Aborted

FILLING: 3:07
HEATING: 16:46
STERILIZING: 4:00
VENTING: 2:53
DRYING: 20:30
TOTAL CYCLE: 00:47:16

10/13/2021 01:33 PM
ACCEPTED: X
REJECTED:
Operator: U01