
Midmark M11 - v1.0.0
Total Cycles: 3930

Name: RITTER 1
Sterilizer ID: V2178198
10/20/2023 01:13 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 110.6 F 0.0
2:00 113.2 F 0.0
4:00 127.6 F 1.2
6:00 163.2 F 4.2
8:00 190.8 F 7.8
10:00 199.0 F 12.0
12:00 216.0 F 4.8
14:00 225.5 F 8.8
16:00 236.6 F 13.1
18:00 235.8 F 9.6
20:00 239.0 F 9.8
22:00 245.6 F 13.2
24:00 256.9 F 19.2
26:00 267.3 F 25.6

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.5
1:00 272.4 F 29.2
2:00 271.5 F 28.4
3:00 272.4 F 29.2
4:00 271.9 F 28.6

Min 271.1 F 28.3
Max 272.5 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.7 F 0.0
5:00 226.6 F 0.0
10:00 222.2 F 0.0
15:00 223.2 F 0.0
20:00 226.6 F 0.0
25:00 224.7 F 0.0
30:00 228.5 F 0.0

Drying Complete

FILLING: 0:56
HEATING: 27:22
STERILIZING: 4:00
VENTING: 3:38
DRYING: 30:00
TOTAL CYCLE: 01:05:56
CYCLE COMPLETE

10/20/2023 02:18 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____