

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 000012  
9:57 19/12/2019  
WRAPPED  
135 C for 10 min.  
CYCLE START 0:00  
CYCLE ABORTED 3:06  
CYCLE FAULT 15  
NOT STERILE  
END TIME 3:06  
-----

Digital Signature #  
FF2FF6160998DE8B  
-----