
Midmark M11 - v1.0.5
Total Cycles: 1289

Name: RITTER2
Sterilizer ID: V2313228
07/13/2022 10:37 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 179.7 F 0.0
2:00 182.5 F 0.2
4:00 199.4 F 3.2
6:00 220.2 F 8.0
8:00 230.3 F 13.3
10:00 233.0 F 9.9
12:00 238.2 F 9.0
14:00 242.6 F 11.9
16:00 245.5 F 13.5
18:00 258.0 F 20.2
20:00 268.9 F 27.1

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.0
1:00 272.5 F 29.6
2:00 271.5 F 28.9
3:00 272.5 F 29.7
4:00 271.8 F 29.0

Min 271.4 F 28.8
Max 272.6 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.8 F 0.0
5:00 216.5 F 0.0
10:00 221.4 F 0.0
15:00 220.7 F 0.0
20:00 219.3 F 0.0
25:00 218.1 F 0.0
30:00 216.1 F 0.0
Drying Complete

FILLING: 1:06
HEATING: 21:11
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:59:26
CYCLE COMPLETE

07/13/2022 11:36 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____