
Midmark M11 - v1.0.5
Total Cycles: 1778

Name: RITTER2
Sterilizer ID: V2313228
11/28/2022 01:29 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 199.3 F 0.0
2:00 195.7 F 4.0
4:00 205.8 F 9.7
6:00 224.9 F 7.5
8:00 235.8 F 12.1
10:00 237.5 F 10.6
12:00 237.6 F 9.9
14:00 252.0 F 17.2
16:00 264.5 F 24.6

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.4
1:00 272.6 F 29.7
2:00 272.3 F 30.0
3:00 272.3 F 29.6
4:00 272.4 F 30.1

Min 271.7 F 29.2
Max 273.2 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.6 F 0.0
5:00 236.3 F 0.0
10:00 237.0 F 0.0
15:00 225.5 F 0.0
20:00 230.1 F 0.0
25:00 221.9 F 0.0
30:00 219.1 F 0.0
Drying Complete

FILLING: 2:35
HEATING: 17:48
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:57:36
CYCLE COMPLETE

11/28/2022 02:26 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____