

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 001735  
12:59 15/04/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:38  
PRESSURIZING 2:45  
STERILIZING 3:36  
136.0 C 220kPa 3:36  
136.1 C 224kPa 4:06  
136.3 C 224kPa 4:36  
136.4 C 226kPa 5:06  
136.3 C 225kPa 5:36  
136.3 C 225kPa 6:06  
136.3 C 224kPa 6:36  
136.4 C 224kPa 7:06  
136.4 C 224kPa 7:06  
VENTING 7:06  
AIR DRYING 7:58  
CYCLE COMPLETE 8:10  
-----

Digital Signature #  
247EB7B41214692F  
-----