
Midmark M11 - v1.0.0
Total Cycles: 1477

Name: RITTER 1
Sterilizer ID: V2178198
12/20/2021 02:40 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 169.9 F 0.0
2:00 171.3 F 2.8
4:00 190.3 F 6.9
6:00 211.8 F 12.4
8:00 223.0 F 6.6
10:00 236.5 F 11.8
12:00 235.5 F 8.9
14:00 239.9 F 9.8
16:00 247.7 F 14.0
18:00 260.1 F 20.8
20:00 271.1 F 28.0

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.3
1:00 272.2 F 28.9
2:00 272.2 F 28.6
3:00 271.5 F 28.4
4:00 272.7 F 29.1

Min 271.4 F 28.1
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.0 F 0.0
5:00 214.0 F 0.0
10:00 216.6 F 0.0
15:00 207.7 F 0.0
20:00 196.9 F 0.0
25:00 192.1 F 0.0
30:00 189.6 F 0.0
Drying Complete

FILLING: 1:25
HEATING: 20:35
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 00:59:09
CYCLE COMPLETE

12/20/2021 03:41 PM
ACCEPTED: X
REJECTED:
Operator: U02