
Midmark M11 - v1.0.0
Total Cycles: 5362

Name: RITTER 1
Sterilizer ID: V2178198
02/25/2025 09:58 AM

BEGIN UNWRAPPED CYCLE
Temp: 270 Degrees F
Time: 3 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.2 F 0.0
2:00 77.0 F 0.4
4:00 91.8 F 1.9
6:00 123.6 F 4.3
8:00 152.8 F 7.1
10:00 173.0 F 10.5
12:00 198.6 F 5.7
14:00 207.2 F 4.1
16:00 218.4 F 7.3
18:00 228.6 F 10.9
20:00 227.3 F 6.8
22:00 238.5 F 10.7
24:00 234.5 F 7.9
26:00 247.5 F 14.0
28:00 258.6 F 19.9
30:00 268.6 F 26.3

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.4
1:00 272.6 F 28.9
2:00 271.8 F 28.5
3:00 272.4 F 28.9

Min 271.6 F 28.2
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 179.6 F 0.0
5:00 171.9 F 0.0
10:00 166.4 F 0.0
15:00 162.4 F 0.0
20:00 161.3 F 0.0
25:00 162.5 F 0.0
30:00 165.6 F 0.0
Drying Complete

FILLING: 0:47
HEATING: 31:14
STERILIZING: 3:00
VENTING: 3:37
DRYING: 30:00
TOTAL CYCLE: 01:08:38
CYCLE COMPLETE

02/25/2025 11:06 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____