
Midmark M11 - v1.0.0
Total Cycles: 1808

Name: RITTER 1
Sterilizer ID: V2178198
04/11/2022 09:54 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 190.6 F 0.0
2:00 181.6 F 3.4
4:00 184.3 F 7.9
6:00 204.8 F 12.9
8:00 224.9 F 7.1
10:00 236.6 F 12.4
12:00 236.0 F 9.1
14:00 238.6 F 9.5
16:00 246.9 F 13.5
18:00 259.0 F 20.1
20:00 269.7 F 27.0

STERILIZING
mm:ss Degrees PSI
0:00 272.1 F 28.5
1:00 272.7 F 29.0
2:00 271.4 F 28.6
3:00 272.6 F 28.8
4:00 271.9 F 28.9

Min 271.4 F 28.4
Max 272.8 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.5 F 0.0
5:00 225.1 F 0.0
10:00 222.2 F 0.0
15:00 233.9 F 0.0
20:00 234.3 F 0.0
25:00 232.3 F 0.0
30:00 226.2 F 0.0
Drying Complete

FILLING: 1:10
HEATING: 21:34
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 00:59:51
CYCLE COMPLETE

04/11/2022 10:53 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____