

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3067

Name: RITTER2  
Sterilizer ID: V2313228  
10/30/2023 04:40 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 163.7 F 0.0  
2:00 167.9 F 3.3  
4:00 195.5 F 9.1  
6:00 215.3 F 7.2  
8:00 229.6 F 11.4  
10:00 234.7 F 9.8  
12:00 235.7 F 8.7  
14:00 251.3 F 16.5  
16:00 264.3 F 24.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 29.1  
1:00 272.4 F 29.6  
2:00 271.4 F 29.4  
3:00 272.4 F 29.6  
4:00 271.2 F 29.2  
  
Min 271.2 F 29.0  
Max 272.7 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 199.3 F 0.0  
5:00 212.2 F 0.0  
10:00 218.0 F 0.0  
15:00 216.7 F 0.0  
20:00 207.5 F 0.0  
25:00 201.7 F 0.0  
30:00 200.0 F 0.0  
Drying Complete

FILLING: 1:60  
HEATING: 17:50  
STERILIZING: 4:00  
VENTING: 3:17  
DRYING: 30:00  
TOTAL CYCLE: 00:57:07  
CYCLE COMPLETE

10/30/2023 05:37 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_