
Midmark M11 - v1.0.0
Total Cycles: 1753

Name: RITTER 1
Sterilizer ID: V2178198
03/28/2022 11:07 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 173.8 F 0.0
2:00 171.3 F 2.8
4:00 183.7 F 6.7
6:00 212.2 F 11.9
8:00 220.8 F 6.8
10:00 234.0 F 12.2
12:00 234.7 F 9.2
14:00 238.7 F 9.4
16:00 247.1 F 14.0
18:00 259.5 F 20.7
20:00 270.5 F 27.8

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.5
1:00 272.5 F 29.2
2:00 271.5 F 28.5
3:00 272.0 F 29.1
4:00 272.4 F 29.1

Min 271.2 F 28.4
Max 272.6 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.5 F 0.0
5:00 216.0 F 0.0
10:00 235.4 F 0.0
15:00 228.2 F 0.0
20:00 229.3 F 0.0
25:00 227.8 F 0.0
30:00 226.2 F 0.0
Drying Complete

FILLING: 1:10
HEATING: 20:51
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 00:59:07
CYCLE COMPLETE

03/28/2022 12:06 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____