
Midmark M11 - v1.0.0
Total Cycles: 3310

Name: RITTER 1
Sterilizer ID: V2178198
05/31/2023 08:35 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 176.1 F 0.0
2:00 172.2 F 0.0
4:00 176.5 F 0.5
6:00 201.6 F 5.0
8:00 218.2 F 10.1
10:00 226.6 F 6.7
12:00 236.9 F 11.0
14:00 234.2 F 8.2
16:00 241.8 F 10.7
18:00 246.9 F 13.5
20:00 259.1 F 20.2
22:00 270.1 F 27.2

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.2
1:00 272.5 F 28.9
2:00 271.9 F 28.3
3:00 271.6 F 28.4
4:00 272.5 F 28.9

Min 271.3 F 28.0
Max 272.6 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.6 F 0.0
5:00 211.5 F 0.0
10:00 211.2 F 0.0
15:00 206.9 F 0.0
20:00 198.5 F 0.0
25:00 200.8 F 0.0
30:00 202.5 F 0.0
Drying Complete

FILLING: 0:48
HEATING: 22:56
STERILIZING: 4:00
VENTING: 3:20
DRYING: 30:00
TOTAL CYCLE: 01:01:04
CYCLE COMPLETE

05/31/2023 09:36 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____