

-----  
Midmark M11 - v1.0.5  
Total Cycles: 19

Name: RITTER2  
Sterilizer ID: V2313228  
04/26/2021 12:47 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 168.3 F 0.0  
2:00 174.3 F 3.9  
4:00 199.2 F 9.6  
6:00 217.2 F 6.6  
8:00 231.5 F 11.6  
10:00 236.2 F 10.0  
12:00 237.1 F 9.3  
14:00 252.3 F 17.0  
16:00 264.8 F 24.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.1  
1:00 272.6 F 29.6  
2:00 271.7 F 29.2  
3:00 272.7 F 29.6  
4:00 271.7 F 28.9  
  
Min 271.5 F 28.9  
Max 272.9 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 187.3 F 0.0  
5:00 234.3 F 0.0  
10:00 222.8 F 0.0  
15:00 227.2 F 0.0  
20:00 192.8 F 0.0  
25:00 198.9 F 0.0  
Dry Cycle Aborted

FILLING: 2:26  
HEATING: 17:48  
STERILIZING: 4:00  
VENTING: 3:07  
DRYING: 28:02  
TOTAL CYCLE: 00:55:23

04/26/2021 01:42 PM  
ACCEPTED: X  
REJECTED:  
Operator: U01