
Midmark M11 - v1.0.5
Total Cycles: 915

Name: RITTER2
Sterilizer ID: V2313228
03/28/2022 12:42 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 132.4 F 0.0
2:00 131.9 F 0.0
4:00 156.4 F 0.2
6:00 195.3 F 1.2
8:00 211.7 F 5.6
10:00 228.0 F 10.2
12:00 231.0 F 8.2
14:00 240.7 F 12.3
16:00 239.6 F 10.8
18:00 238.2 F 9.9
20:00 252.7 F 16.9
22:00 264.6 F 24.0

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 28.7
1:00 271.9 F 29.3
2:00 272.5 F 29.2
3:00 271.3 F 28.6
4:00 272.2 F 29.6

Min 271.2 F 28.5
Max 272.7 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.7 F 0.0
5:00 220.9 F 0.0
10:00 226.1 F 0.0
15:00 230.7 F 0.0
20:00 225.9 F 0.0
25:00 226.7 F 0.0
30:00 226.8 F 0.0
Drying Complete

FILLING: 1:03
HEATING: 23:46
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 01:01:60
CYCLE COMPLETE

03/28/2022 01:43 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____