
Midmark M11 - v1.0.5
Total Cycles: 613

Name: RITTER2
Sterilizer ID: V2313228
12/08/2021 01:10 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 156.4 F 0.0
2:00 153.1 F 3.0
4:00 174.7 F 7.0
6:00 197.1 F 12.0
8:00 221.9 F 5.9
10:00 234.2 F 11.1
12:00 233.0 F 8.4
14:00 241.2 F 11.0
16:00 245.9 F 13.6
18:00 258.5 F 20.4
20:00 269.7 F 27.7

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.9
1:00 272.7 F 29.5
2:00 271.4 F 28.8
3:00 272.9 F 29.7
4:00 272.0 F 29.0

Min 271.3 F 28.8
Max 272.9 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 182.4 F 0.0
5:00 238.9 F 0.0
10:00 200.4 F 0.0
15:00 190.0 F 0.0
20:00 183.6 F 0.0
25:00 196.6 F 0.0
30:00 181.5 F 0.0
Drying Complete

FILLING: 1:08
HEATING: 20:57
STERILIZING: 4:00
VENTING: 2:60
DRYING: 30:00
TOTAL CYCLE: 00:59:05
CYCLE COMPLETE

12/08/2021 02:15 PM
ACCEPTED: X
REJECTED:
Operator: U02