
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002504
14:05 10/02/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:30
PRESSURIZING 2:36
STERILIZING 3:37
136.0 C 220kPa 3:37
136.4 C 226kPa 4:07
136.2 C 223kPa 4:37
136.3 C 225kPa 5:07
136.2 C 225kPa 5:37
136.4 C 227kPa 6:07
136.4 C 226kPa 6:37
136.4 C 224kPa 7:07
136.4 C 224kPa 7:07
VENTING 7:07
AIR DRYING 8:07
CYCLE COMPLETE 34:59

Digital Signature #
992EB54458928128
