
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 001237
8:26 23/02/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:45
PRESSURIZING 3:48
STERILIZING 4:59
136.0 C 220kPa 4:59
136.4 C 221kPa 5:29
136.1 C 221kPa 5:59
136.4 C 224kPa 6:29
136.3 C 223kPa 6:59
136.1 C 222kPa 7:29
136.3 C 222kPa 7:59
136.1 C 221kPa 8:29
136.1 C 221kPa 8:29
VENTING 8:29
AIR DRYING 9:23
CYCLE COMPLETE 9:24

Digital Signature #
00157105E80614B2
