
Midmark M11 - v1.0.5
Total Cycles: 3984

Name: RITTER2
Sterilizer ID: V2313228
10/01/2024 02:03 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 123.0 F 0.0
2:00 124.0 F 0.5
4:00 139.8 F 3.1
6:00 165.8 F 6.3
8:00 185.9 F 9.8
10:00 199.7 F 13.5
12:00 215.8 F 6.8
14:00 225.0 F 11.0
16:00 229.1 F 7.4
18:00 239.8 F 12.6
20:00 239.7 F 10.9
22:00 251.3 F 16.6
24:00 261.1 F 22.5
26:00 270.8 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.6
1:00 272.3 F 29.9
2:00 271.9 F 30.1
3:00 271.3 F 29.7
4:00 271.7 F 29.7

Min 271.3 F 29.5
Max 272.4 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 193.4 F 0.0
5:00 220.4 F 0.0
10:00 214.2 F 0.0
15:00 198.4 F 0.0
20:00 191.2 F 0.0
25:00 186.8 F 0.0
30:00 183.4 F 0.0
Drying Complete

FILLING: 0:59
HEATING: 26:37
STERILIZING: 4:00
VENTING: 3:51
DRYING: 30:00
TOTAL CYCLE: 01:05:26
CYCLE COMPLETE

10/01/2024 03:08 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____