

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2748

Name: RITTER 1  
Sterilizer ID: V2178198  
01/06/2023 03:56 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 86.3 F 0.0  
2:00 94.8 F 0.0  
4:00 112.2 F 0.0  
6:00 145.8 F 0.0  
8:00 178.3 F 0.7  
10:00 193.1 F 4.5  
12:00 206.1 F 8.1  
14:00 218.1 F 11.7  
16:00 223.1 F 5.6  
18:00 234.9 F 10.2  
20:00 236.5 F 8.9  
22:00 244.3 F 12.6  
24:00 241.9 F 11.0  
26:00 254.8 F 17.5  
28:00 266.1 F 24.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 27.9  
1:00 272.6 F 28.5  
2:00 271.5 F 28.1  
3:00 272.6 F 28.5  
4:00 271.5 F 27.9

Min 271.4 F 27.8  
Max 272.8 F 28.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 174.8 F 0.0  
5:00 204.4 F 0.0  
10:00 210.2 F 0.0  
15:00 215.5 F 0.0  
20:00 216.6 F 0.0  
25:00 214.7 F 0.0  
30:00 213.5 F 0.0  
Drying Complete

FILLING: 0:58  
HEATING: 29:39  
STERILIZING: 4:00  
VENTING: 3:07  
DRYING: 30:00  
TOTAL CYCLE: 01:07:44  
CYCLE COMPLETE

01/06/2023 05:03 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_