

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 007991  
12:56 23/10/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:39  
PRESSURIZING 2:42  
STERILIZING 3:46  
136.5 C 223kPa 3:46  
136.4 C 224kPa 4:16  
136.3 C 222kPa 4:46  
136.4 C 224kPa 5:16  
136.4 C 221kPa 5:46  
136.4 C 223kPa 6:16  
136.4 C 224kPa 6:46  
136.1 C 221kPa 7:16  
136.1 C 221kPa 7:16  
VENTING 7:16  
AIR DRYING 8:05  
CYCLE COMPLETE 25:55  
-----

Digital Signature #  
671AE62202E4F97A  
-----