
Midmark M11 - v1.0.5
Total Cycles: 1776

Name: RITTER2
Sterilizer ID: V2313228
11/28/2022 11:30 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 205.4 F 0.0
2:00 201.5 F 3.6
4:00 203.5 F 9.1
6:00 228.8 F 8.2
8:00 235.9 F 12.1
10:00 239.4 F 10.6
12:00 237.4 F 9.8
14:00 252.9 F 17.1
16:00 264.5 F 24.4

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.2
1:00 272.2 F 29.5
2:00 271.8 F 29.8
3:00 271.7 F 29.4
4:00 272.0 F 29.9

Min 271.2 F 29.0
Max 272.7 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.2 F 0.0
5:00 238.8 F 0.0
10:00 229.1 F 0.0
15:00 229.2 F 0.0
20:00 226.6 F 0.0
25:00 225.5 F 0.0
30:00 223.6 F 0.0
Drying Complete

FILLING: 2:32
HEATING: 17:48
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:57:33
CYCLE COMPLETE

11/28/2022 12:27 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____