

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1619

Name: RITTER2  
Sterilizer ID: V2313228  
10/18/2022 02:57 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 196.9 F 0.0  
2:00 192.2 F 4.0  
4:00 197.9 F 9.5  
6:00 215.5 F 6.9  
8:00 232.4 F 11.5  
10:00 235.6 F 10.0  
12:00 236.7 F 9.3  
14:00 251.5 F 16.9  
16:00 264.2 F 24.5

STERILIZING  
mm:ss Degrees PSI  
0:00 272.0 F 29.4  
1:00 272.7 F 30.1  
2:00 272.4 F 29.7  
3:00 271.9 F 29.6  
4:00 273.0 F 30.2  
  
Min 271.7 F 29.3  
Max 273.1 F 30.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 206.0 F 0.0  
5:00 209.2 F 0.0  
10:00 219.4 F 0.0  
15:00 204.6 F 0.0  
20:00 200.5 F 0.0  
25:00 198.6 F 0.0  
30:00 196.1 F 0.0  
Drying Complete

FILLING: 4:09  
HEATING: 17:48  
STERILIZING: 4:00  
VENTING: 3:17  
DRYING: 30:00  
TOTAL CYCLE: 00:59:14  
CYCLE COMPLETE

10/18/2022 03:56 PM  
ACCEPTED:  
REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_