
Midmark M11 - v1.0.0
Total Cycles: 1075

Name: RITTER 1
Sterilizer ID: V2178198
08/05/2021 12:27 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 110.4 F 0.0
2:00 113.5 F 0.0
4:00 131.4 F 0.0
6:00 168.1 F 0.2
8:00 195.0 F 3.7
10:00 210.7 F 7.3
12:00 222.8 F 11.1
14:00 224.6 F 5.8
16:00 236.6 F 10.7
18:00 233.2 F 7.5
20:00 246.2 F 13.5
22:00 244.1 F 12.1
24:00 256.5 F 18.5
26:00 267.5 F 25.2

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.0
1:00 272.5 F 28.6
2:00 271.4 F 28.0
3:00 272.6 F 28.7
4:00 271.6 F 27.9

Min 271.4 F 27.9
Max 272.7 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.1 F 0.0
5:00 203.7 F 0.0
10:00 194.0 F 0.0
15:00 195.4 F 0.0
20:00 192.1 F 0.0
25:00 188.6 F 0.0
30:00 187.5 F 0.0
Drying Complete

FILLING: 1:03
HEATING: 27:25
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 01:05:35
CYCLE COMPLETE

08/05/2021 02:15 PM
ACCEPTED: X
REJECTED:
Operator: U02