
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 007476
10:25 03/06/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:04
PRESSURIZING 3:07
STERILIZING 4:09
136.0 C 220kPa 4:09
136.4 C 225kPa 4:39
136.4 C 225kPa 5:09
136.4 C 224kPa 5:39
136.4 C 223kPa 6:09
136.2 C 224kPa 6:39
136.4 C 223kPa 7:09
136.0 C 222kPa 7:39
136.0 C 222kPa 7:39
VENTING 7:39
AIR DRYING 8:30
CYCLE COMPLETE 8:39

Digital Signature #
DAD05837A15ECA9F
