
Midmark M11 - v1.0.0
Total Cycles: 2266

Name: RITTER 1
Sterilizer ID: V2178198
08/30/2022 12:17 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 171.3 F 0.0
2:00 173.5 F 1.0
4:00 185.6 F 4.2
6:00 207.3 F 9.1
8:00 220.4 F 7.2
10:00 233.0 F 9.4
12:00 236.4 F 9.7
14:00 243.4 F 12.3
16:00 241.7 F 11.2
18:00 255.6 F 18.3
20:00 267.2 F 25.3

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.3
1:00 272.4 F 28.7
2:00 272.1 F 28.7
3:00 272.3 F 28.6
4:00 272.0 F 28.7

Min 271.6 F 28.1
Max 272.8 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.7 F 0.0
5:00 224.5 F 0.0
10:00 217.6 F 0.0
15:00 219.8 F 0.0
20:00 224.3 F 0.0
25:00 215.1 F 0.0
30:00 213.1 F 0.0
Drying Complete

FILLING: 1:28
HEATING: 21:22
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 01:00:05
CYCLE COMPLETE

08/30/2022 01:17 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____