

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4461

Name: RITTER2  
Sterilizer ID: V2313228  
03/26/2025 03:03 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 129.0 F 0.0  
2:00 129.2 F 1.5  
4:00 134.7 F 4.4  
6:00 162.6 F 8.4  
8:00 182.7 F 12.9  
10:00 212.1 F 5.7  
12:00 225.2 F 10.0  
14:00 233.4 F 9.1  
16:00 238.4 F 12.0  
18:00 237.9 F 10.5  
20:00 250.4 F 16.7  
22:00 261.3 F 23.2  
24:00 271.9 F 30.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 29.8  
1:00 272.5 F 30.5  
2:00 271.5 F 29.8  
3:00 271.5 F 30.5  
4:00 272.3 F 30.3

Min 271.1 F 29.7  
Max 272.6 F 30.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 207.3 F 0.0  
5:00 194.4 F 0.0  
10:00 204.8 F 0.0  
15:00 203.2 F 0.0  
20:00 202.0 F 0.0  
25:00 199.9 F 0.0  
30:00 198.5 F 0.0  
Drying Complete

FILLING: 1:15  
HEATING: 24:32  
STERILIZING: 4:00  
VENTING: 3:56  
DRYING: 30:00  
TOTAL CYCLE: 01:03:44  
CYCLE COMPLETE

03/26/2025 04:06 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_