
Midmark M11 - v1.0.0
Total Cycles: 651

Name: MIDMARK
Sterilizer ID: V2178198
02/02/2021 02:35 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 165.7 F 0.0
2:00 165.0 F 2.8
4:00 176.9 F 7.1
6:00 203.1 F 12.8
8:00 222.5 F 7.9
10:00 241.5 F 12.7
12:00 240.0 F 10.9
14:00 236.8 F 9.0
16:00 251.5 F 16.1
18:00 263.9 F 23.2
20:00 271.3 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.8
1:00 272.4 F 28.9
2:00 272.6 F 29.3
3:00 272.4 F 29.5
4:00 271.8 F 28.7

Min 271.7 F 28.7
Max 272.8 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 214.6 F 0.0
5:00 232.9 F 0.0
10:00 229.4 F 0.0
15:00 226.9 F 0.0
20:00 226.2 F 0.0
Dry Cycle Aborted

FILLING: 1:21
HEATING: 20:35
STERILIZING: 4:00
VENTING: 2:59
DRYING: 23:36
TOTAL CYCLE: 00:52:30

02/02/2021 03:27 PM
ACCEPTED:
REJECTED:

Reason: _____
Operator: _____