

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
1.9uS / 1.2ppm  
CYCLE NUMBER 001020  
8:44 29/03/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:26  
PRESSURIZING 2:30  
STERILIZING 3:23  
136.5 C 225kPa 3:23  
136.4 C 225kPa 3:53  
136.2 C 224kPa 4:23  
136.3 C 224kPa 4:53  
135.9 C 220kPa 5:23  
136.2 C 223kPa 5:53  
136.2 C 223kPa 6:23  
136.2 C 223kPa 6:53  
136.2 C 223kPa 6:53  
VENTING 6:53  
AIR DRYING 7:39  
CYCLE COMPLETE 7:46  
-----

Digital Signature #  
E388435147A414DA  
-----