
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 002187
7:55 01/11/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:23
PRESSURIZING 2:30
STERILIZING 3:30
136.0 C 220kPa 3:30
136.0 C 221kPa 4:00
136.2 C 222kPa 4:30
136.4 C 224kPa 5:00
136.2 C 219kPa 5:30
136.4 C 223kPa 6:00
136.4 C 224kPa 6:30
136.3 C 222kPa 7:00
136.3 C 222kPa 7:00
VENTING 7:01
AIR DRYING 7:51
CYCLE COMPLETE 14:06

Digital Signature #
FE3D506F020F0551
