
Midmark M11 - v1.0.0
Total Cycles: 5194

Name: RITTER 1
Sterilizer ID: V2178198
12/16/2024 03:19 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 99.7 F 0.0
2:00 101.9 F 0.0
4:00 128.5 F 0.0
6:00 146.6 F 0.1
8:00 175.5 F 0.3
10:00 195.7 F 2.4
12:00 217.3 F 5.6
14:00 224.0 F 9.3
16:00 233.4 F 13.5
18:00 235.0 F 9.3
20:00 241.7 F 10.8
22:00 244.0 F 12.2
24:00 241.4 F 10.8
26:00 253.9 F 17.0
28:00 264.3 F 23.1
30:00 272.2 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 27.9
1:00 271.7 F 28.2
2:00 272.6 F 28.7
3:00 272.1 F 28.2
4:00 271.5 F 27.8

Min 271.4 F 27.8
Max 272.6 F 28.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 186.2 F 0.0
5:00 209.2 F 0.0
10:00 205.4 F 0.0
15:00 192.4 F 0.0
20:00 205.3 F 0.0
25:00 221.0 F 0.0
30:00 231.0 F 0.0
Drying Complete

FILLING: 0:45
HEATING: 30:06
STERILIZING: 4:00
VENTING: 3:50
DRYING: 30:00
TOTAL CYCLE: 01:08:41
CYCLE COMPLETE

12/16/2024 04:27 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____