
Midmark M11 - v1.0.0
Total Cycles: 1478

Name: RITTER 1
Sterilizer ID: V2178198
12/20/2021 03:45 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 170.8 F 0.0
2:00 180.7 F 3.1
4:00 201.9 F 7.5
6:00 222.9 F 12.8
8:00 228.9 F 7.5
10:00 240.1 F 12.8
12:00 239.0 F 10.4
14:00 235.9 F 8.6
16:00 250.9 F 15.6
18:00 262.7 F 22.4
20:00 272.5 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.2
1:00 272.7 F 28.8
2:00 271.5 F 28.1
3:00 272.7 F 29.0
4:00 271.9 F 28.4

Min 271.4 F 28.1
Max 272.8 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.5 F 0.0
5:00 224.7 F 0.0
10:00 203.7 F 0.0
15:00 203.7 F 0.0
20:00 200.2 F 0.0
25:00 202.0 F 0.0
30:00 201.2 F 0.0
Drying Complete

FILLING: 1:40
HEATING: 20:14
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 00:59:10
CYCLE COMPLETE

12/20/2021 04:44 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____