
Midmark M11 - v1.0.5
Total Cycles: 1434

Name: RITTER2
Sterilizer ID: V2313228
08/30/2022 12:36 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 137.6 F 0.0
2:00 140.8 F 0.0
4:00 159.0 F 2.3
6:00 191.0 F 5.8
8:00 207.1 F 10.4
10:00 217.2 F 6.0
12:00 231.5 F 11.0
14:00 232.4 F 8.7
16:00 240.1 F 10.4
18:00 247.1 F 14.6
20:00 259.8 F 21.7
22:00 271.3 F 29.2

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.4
1:00 272.9 F 30.0
2:00 271.9 F 29.4
3:00 272.9 F 30.0
4:00 272.1 F 29.3

Min 271.8 F 29.2
Max 273.1 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 193.7 F 0.0
5:00 201.9 F 0.0
10:00 208.3 F 0.0
15:00 213.3 F 0.0
20:00 217.1 F 0.0
25:00 213.7 F 0.0
30:00 205.4 F 0.0
Drying Complete

FILLING: 1:18
HEATING: 22:40
STERILIZING: 4:00
VENTING: 3:05
DRYING: 30:00
TOTAL CYCLE: 01:01:03
CYCLE COMPLETE

08/30/2022 01:37 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____