
Midmark M11 - v1.0.5
Total Cycles: 2759

Name: RITTER2
Sterilizer ID: V2313228
08/14/2023 12:05 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 184.6 F 0.0
2:00 183.5 F 0.0
4:00 199.6 F 2.2
6:00 214.5 F 6.2
8:00 226.6 F 10.7
10:00 229.3 F 7.4
12:00 241.1 F 12.5
14:00 240.3 F 11.1
16:00 239.5 F 10.4
18:00 252.7 F 16.9
20:00 263.7 F 23.4
22:00 272.2 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.7
1:00 272.4 F 29.0
2:00 272.6 F 29.5
3:00 271.6 F 28.6
4:00 272.6 F 29.2

Min 271.5 F 28.5
Max 272.8 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.6 F 0.0
5:00 218.0 F 0.0
10:00 220.5 F 0.0
15:00 225.5 F 0.0
20:00 223.0 F 0.0
25:00 221.2 F 0.0
30:00 218.8 F 0.0
Drying Complete

FILLING: 1:46
HEATING: 22:08
STERILIZING: 4:00
VENTING: 3:29
DRYING: 30:00
TOTAL CYCLE: 01:01:23
CYCLE COMPLETE

08/14/2023 01:06 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____