

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2654

Name: RITTER2  
Sterilizer ID: V2313228  
07/17/2023 12:21 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 169.7 F 0.0  
2:00 170.1 F 2.7  
4:00 184.5 F 6.6  
6:00 205.9 F 11.6  
8:00 220.0 F 6.3  
10:00 231.9 F 11.3  
12:00 233.2 F 8.7  
14:00 239.9 F 10.7  
16:00 246.0 F 13.7  
18:00 258.0 F 20.3  
20:00 269.0 F 27.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 29.1  
1:00 272.3 F 29.3  
2:00 272.5 F 29.8  
3:00 271.8 F 29.0  
4:00 272.6 F 29.7

Min 271.6 F 28.9  
Max 272.7 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 200.5 F 0.0  
5:00 220.3 F 0.0  
10:00 219.5 F 0.0  
15:00 217.2 F 0.0  
20:00 218.3 F 0.0  
25:00 217.3 F 0.0  
30:00 213.2 F 0.0  
Drying Complete

FILLING: 1:40  
HEATING: 21:04  
STERILIZING: 4:00  
VENTING: 3:16  
DRYING: 30:00  
TOTAL CYCLE: 00:59:60  
CYCLE COMPLETE

07/17/2023 01:20 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_