
Midmark M11 - v1.0.5
Total Cycles: 1755

Name: RITTER2
Sterilizer ID: V2313228
11/21/2022 04:43 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 211.9 F 0.0
2:00 196.3 F 3.9
4:00 214.2 F 9.8
6:00 227.6 F 7.1
8:00 237.8 F 12.9
10:00 241.8 F 11.7
12:00 241.7 F 11.4
14:00 255.9 F 18.8
16:00 268.1 F 26.3

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.7
1:00 272.3 F 29.1
2:00 271.7 F 29.3
3:00 272.0 F 29.0
4:00 271.6 F 29.2

Min 271.2 F 28.6
Max 272.6 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 198.3 F 0.0
5:00 236.3 F 0.0
10:00 238.1 F 0.0
15:00 228.8 F 0.0
20:00 228.4 F 0.0
25:00 228.6 F 0.0
30:00 227.5 F 0.0
Drying Complete

FILLING: 2:28
HEATING: 17:10
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 00:56:53
CYCLE COMPLETE

11/21/2022 05:39 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____