
Surrey, BC
Canada

STATIM 5000 S501R711
SN 511018F00076
Unit # : 000
WATER QUALITY
2.3uS / 1.4ppm
CYCLE NUMBER 000016
15:20 13/01/2020
WRAPPED
134 C for 6 min.
CYCLE START 0:00
CONDITIONING 2:11
PRESSURIZING 13:31
STERILIZING 15:11
136.4 C 227kPa 15:11
136.4 C 227kPa 15:41
136.5 C 227kPa 16:11
136.4 C 227kPa 16:41
136.4 C 226kPa 17:11
136.4 C 225kPa 17:42
136.4 C 226kPa 18:12
136.4 C 226kPa 18:42
136.4 C 227kPa 19:12
136.4 C 225kPa 19:42
136.4 C 227kPa 20:12
136.3 C 225kPa 20:42
136.4 C 227kPa 21:11
VENTING 21:11
AIR DRYING 22:27
CYCLE COMPLETE 24:28

Digital Signature #
A77C4CCAF1C08F6C
