
Midmark M11 - v1.0.0
Total Cycles: 1067

Name: RITTER 1
Sterilizer ID: V2178198
07/28/2021 12:43 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 175.1 F 0.0
2:00 180.8 F 3.2
4:00 185.4 F 7.4
6:00 200.7 F 12.1
8:00 212.9 F 7.5
10:00 228.6 F 12.7
12:00 232.0 F 10.2
14:00 233.0 F 8.5
16:00 246.8 F 14.9
18:00 258.6 F 21.3
20:00 269.3 F 28.0

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.5
1:00 272.8 F 30.2
2:00 272.2 F 29.6
3:00 272.4 F 30.0
4:00 272.7 F 30.1

Min 271.7 F 29.4
Max 272.9 F 30.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.8 F 0.0
5:00 213.5 F 0.0
10:00 223.0 F 0.0
15:00 227.0 F 0.0
20:00 231.9 F 0.0
25:00 227.9 F 0.0
30:00 223.2 F 0.0
Drying Complete

FILLING: 1:53
HEATING: 21:04
STERILIZING: 4:00
VENTING: 3:29
DRYING: 30:00
TOTAL CYCLE: 01:00:26
CYCLE COMPLETE

07/28/2021 02:00 PM
ACCEPTED: X
REJECTED:
Operator: U01