
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 005221
9:47 14/03/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:56
PRESSURIZING 2:00
STERILIZING 3:03
136.5 C 220kPa 3:03
136.4 C 221kPa 3:33
136.3 C 224kPa 4:03
136.4 C 223kPa 4:33
136.3 C 223kPa 5:03
136.4 C 223kPa 5:33
136.3 C 223kPa 6:03
136.4 C 222kPa 6:33
136.4 C 222kPa 6:33
VENTING 6:33
AIR DRYING 7:23
CYCLE COMPLETE 13:25

Digital Signature #
DC8B8B6991F4F147
