

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1957

Name: RITTER 1  
Sterilizer ID: V2178198  
05/26/2022 10:52 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 187.7 F 0.0  
2:00 176.4 F 3.4  
4:00 214.5 F 8.0  
6:00 229.2 F 11.6  
8:00 228.4 F 8.1  
10:00 239.2 F 13.2  
12:00 239.0 F 10.3  
14:00 235.7 F 7.9  
16:00 249.7 F 15.0  
18:00 261.6 F 21.7  
20:00 272.4 F 28.8

STERILIZING  
mm:ss Degrees PSI  
0:00 272.4 F 28.6  
1:00 272.6 F 29.0  
2:00 272.7 F 29.3  
3:00 271.7 F 28.7  
4:00 272.4 F 28.8

Min 271.6 F 28.5  
Max 272.9 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 208.4 F 0.0  
5:00 234.1 F 0.0  
10:00 224.8 F 0.0  
15:00 217.2 F 0.0  
20:00 218.1 F 0.0  
25:00 228.5 F 0.0  
30:00 215.6 F 0.0  
Drying Complete

FILLING: 1:52  
HEATING: 20:59  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 30:00  
TOTAL CYCLE: 01:00:03  
CYCLE COMPLETE

05/26/2022 11:55 AM  
ACCEPTED: X  
REJECTED:  
Operator: U02