
Midmark M11 - v1.0.0
Total Cycles: 3359

Name: RITTER 1
Sterilizer ID: V2178198
06/08/2023 01:08 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 194.4 F 0.0
2:00 188.6 F 0.7
4:00 202.2 F 5.4
6:00 218.2 F 9.9
8:00 230.1 F 8.3
10:00 233.3 F 11.0
12:00 234.3 F 8.6
14:00 240.2 F 10.5
16:00 247.2 F 13.8
18:00 259.2 F 20.4
20:00 270.1 F 27.3

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.7
1:00 272.7 F 28.9
2:00 272.5 F 29.3
3:00 272.1 F 28.7
4:00 272.7 F 29.3

Min 271.8 F 28.6
Max 272.9 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.8 F 0.0
5:00 231.0 F 0.0
10:00 230.3 F 0.0
15:00 230.4 F 0.0
20:00 231.2 F 0.0
25:00 229.7 F 0.0
30:00 232.1 F 0.0
Drying Complete

FILLING: 1:41
HEATING: 21:39
STERILIZING: 4:00
VENTING: 3:32
DRYING: 30:00
TOTAL CYCLE: 01:00:52
CYCLE COMPLETE

06/08/2023 02:08 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____